## 

## EC DECLARATION OF CONFORMITY

EUROLATEX SDN. BHD Factory Address: Plot 33, Kuala Ketil Industrial Estate, 09300 Kuala Ketil, Kedah, <u>MALAYSIA</u>

declares that the medical device described hereafter

## Latex Probe covers

have been classified as **Class IIa** device in accordance with Rule 5 of Annex IX and are in conformity with the essential requirement and provision of Medical Device Directive 93/42/EEC as amended by Directive 2007/47/EC and is subject to the procedure set out in Annex II of Directive 93/42/EEC as amended by Directive 2007/47/EC under the supervision of Notified Body (BSI).

EU Representative:

CIRIANO GLOBAL S.L. C/Blancas 4-6, 1 B 50001 Zaragoza, Spain

Date: 03/02/2017

Mr. Subramaniam Managing Director EUROLATEX SDN. BHD